
NOTES:

1. SUBSTRATE:

Calcium Fluoride (CaF2) Vacuum UV Grade

- 2. SURFACE \$2 TO BE PARALLEL WITH \$1 TO <1
- 3. COATING (APPLY ACROSS CLEAR APERTURE) \$1 & \$2: BBAR (2000-5000nm)
- 4. FINE GRIND SURFACE



FOR INFORMATION ONLY: DO NOT MANUFACTURE PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

	S1	\$2
SHAPE	PLANO	PLANO
SURFACE QUALITY	40-20	40-20
CLEAR APERTURE	Ø 18.00	Ø 18.00
BEVEL	PROTECTED AS NEEDED	PROTECTED AS NEEDED

GO ® E	Edmund O	ptics®
---------------	----------	--------

HIRD ANGLE PROJECTION	ϕ	TITLE	20mm Dia. x 2mm Thick, BBAR (2000-5000nm) Coated Calcium Fluoride Window	
ALL DIMS IN	mm	DWG NO	23514	SHEET 8 OF 28